

# CALIBRATION CERTIFICATE

for

## Kustom LaserCam 4 Speed Meter

This Certificate is a Re-Issue and Replaces Certificate Number: 7901

**Customer:** Dorset Police **Serial Number:** LE03365  
**Certificate Number:** 7901 [R]  
**Date of Calibration:** 11 April 2024 **Next Calibration:** 10 April 2025  
**Date of Issue:** 06 June 2024  
**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

### Calibration Results:

Checksum - CD748BAF  
Oscillator Frequency - 19.999809 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
Horizontal - Pass  
Eye Safety - Pass

### Range ± 0.2 m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

### Speed ± 1 MPH

Actual	44 MPH	30 MPH	-21 MPH	52 MPH	-73 MPH
Indicated	44 MPH	30 MPH	-21 MPH	52 MPH	-73 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification  
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

  
L. Korcz  
Technical Signatory

# TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: [info@truelouk.com](mailto:info@truelouk.com)